

Infinity 1030 Series *User Manual*



Contents

Change History	3
How to use this manual	3
Warranties and Liabilities	4
Safety Notices	5
Symbols	5
Safety Notice to Users	5
Certificate of Compliance	6
Technical Specifications	7
Environment	7
Inlet Conditions	7
Outlet Gas	7
General	7
Introduction	8
Unpacking and Installation	8
Fittings Kit	9
General Dimensions	9
Air Connection	11
Commissioning	12
Pressure & Flow Settings	12
Cleaning	12
Maintenance	13
Service Schedule	13
Safely Isolated Condition	13
IMPORTANT DOCUMENTS	14
Warranty Entitlement	14
Supported Maintenance Guide	15
Service Parts List	17

Change History

Rev.	Comment	Name	Date
1	Added 1031 Hi-Flow Information & Content Update	Liam Couttie	24/06/2014
2A	Updated Fittings Kit	Liam Couttie	01/10/2014
2B	Technical Specification Update	Liam Couttie	12/02/2015
3	Technical Specification Update	Liam Couttie	25/05/2016
4	Technical Specification Update	Liam Couttie	31/08/2016
5	Technical Specification Update	Liam Couttie	23/01/2018
6	Service Kit Ordering Update	Liam Couttie	23/03/2018
7	Dimensions Update	Liam Couttie	02/03/2020
8	Declarations Update	Liam Couttie	26/10/2021
9	Declarations Update	Liam Couttie	14/01/2022
10	SMG Guide Added	Cleo Denholm	24/06/2024

How to use this manual

This manual is intended for end users and has been written so that it can either be read as a step by step guide to installation and usage; or as a reference document where you can skip to the relevant information.

Please review each of the following sections carefully.

Thank you for selecting Peak Scientific to meet your Gas Generation needs, should you require any further assistance or support please do not hesitate to contact Peak Scientific or the Peak Partner from which you purchased your Generator.

Warranties and Liabilities

1. The Company warrants that it has title to the Goods.
2. Subject to the provisions of this clause the Company warrants that the Goods shall comply in all material respects with any specification referred to in the Order Confirmation (as the same may be amended) and shall, subject thereto, be free from defects in material and workmanship for the lesser of a period of twelve months from the date of delivery or thirteen months from the date of dispatch from the factory.
3. Save as provided in this clause and except where the Goods are sold to a person dealing as a consumer (within the meaning of the Unfair Contract Terms Act 1977) all warranties, conditions or other terms implied by statute or common law are hereby expressly excluded save to the extent they may not be lawfully excluded. When the Goods are sold to a consumer within the meaning of the Unfair Contract Terms Act 1977 their statutory rights are not affected by the provisions of this clause.
4. In the event of the Customer making a claim in respect of any defect in terms of clause 2 hereof the Customer must.
 - 4.1. Reasonably satisfy the Company that the Goods have been properly installed, commissioned, stored, serviced and used and without prejudice to the generality of the foregoing that any defect is not the direct or indirect result of lack of repair and/or servicing, incorrect repair and/or servicing, use of wrong materials and/or incorrect spare parts
 - 4.2. Allow the company to inspect the Goods and/or any installation and any relevant packaging as and when reasonably required by the Company.
5. Subject to the Company being notified of any defect as is referred to in sub-clause 2 hereof within a reasonable time of it becoming apparent and subject always to the terms of sub-clause 4 hereof, the Company shall, in its option, replace or repair the defective Goods or refund a proportionate part of the Price. The Company shall have no further liability to the Customer (save as mentioned in sub-clause 6 hereof).
6. The Company shall be liable to indemnify the Customer in respect of any claim for death or personal injury to any person in so far as such is attributable to the negligence or breach of duty of the Company or any failure by the Company to comply with the provisions of sub-clause 2 hereof.
7. Save as provided in sub-clause 2 hereof the Company shall not be liable in respect of any claim by the Customer for costs, damages, loss or expenses (whether direct, indirect, consequential or otherwise) or indemnity in any respect howsoever arising including, but not by way of limitation, liability arising in negligence (other than pursuant to clause 6 above) that may be suffered by the Customer or any third party.

Safety Notices

Symbols

This manual uses the following symbols to highlight specific areas important to the safe and proper use of the Generator.



	A WARNING notice denotes a hazard. It calls attention to an operating procedure, process or similar, which if not correctly performed or adhered to, could cause personal injury or in the worst case death. Do not proceed beyond a WARNING notice until the indicated conditions are fully understood or met.
	A CAUTION notice denotes a hazard. It calls attention to an operating procedure, process or similar, which if not correctly performed or adhered to, could cause damage to the Generator or the Application. Do not proceed beyond a CAUTION notice until the indicated conditions are fully understood or met.

Table 1 - Safety Symbols

Safety Notice to Users

These instructions must be read thoroughly and understood before installation and operation of your Peak Scientific Infinity 103*. Use of the Generator in a manner not specified by Peak Scientific MAY impair the SAFETY provided by the equipment.

When handling, operating or carrying out any maintenance, personnel must employ safe engineering practices and observe all relevant local health and safety requirements and regulations. The attention of UK users is drawn to the Health and Safety at Work Act 1974, and the Institute of Electrical Engineers regulations.



Nitrogen is not a poisonous gas, but if the concentration in the inhaled air becomes too high there will be a risk of asphyxiation.

Certificate of Compliance

We Peak Scientific Instruments Ltd.
Of Fountain Crescent, Inchinnan, Renfrewshire, PA4 9RE

Hereby declare that:

Equipment Type: Nitrogen Gas Generator
Model Designator: Infinity 103X

To which this certificate relates is not required to carry a CE or UKCA mark as they fall out with the scope of all relevant EU Directives and UK Regulations, due to their lack of electrical functionality and its operating pressure.

However the following can be stated :

It is manufactured in a facility which is certified to and operates in accordance with the requirements of ISO9001:2015 .

All components used in its manufacture are RoHS III compliant in line with the requirements of EU Directive 2015/863.

Once disconnected from its air supply there are no specific disassembly or material hazards to be considered at end of life.

For and on behalf of Peak Scientific Instruments Ltd.,

Signed:



Name: Fraser Dunn

Position: Design Engineering Manager
Peak Scientific Instruments Ltd,
Inchinnan, Renfrew, Scotland, PA4 9RE, UK.

Date: 14th January 2022

Technical Specifications

Environment

	1031	Hi-Flow	1032	1033	1034
Min/Max Air Operating Temperature	5°C - 30°C / 41°F - 86°F				
Max Altitude	2000 metres				
Maximum Relative Humidity	70%				
Min/Max Storage Temperature	-20°C - 60°C / -4°F - 140°F				

Inlet Conditions

Min/Max Air Inlet Pressure	8.3-10 bar / 120-145 psi				
Minimum Air Inlet Flow (l/min)	115	115	240	340	475

Outlet Gas

Maximum Gas Output Pressure	110 psig				
Maximum Pressure Drop Outlet Inlet	8 psig				
Max Curtain Gas Output Flow	19 l/min	10 l/min	38 l/min	57 l/min	76 l/min
Max Source Gas Output Flow	26 l/min	22 l/min	52 l/min	78 l/min	104 l/min
Max Exhaust Gas Output Flow	25 l/min	10 l/min	50 l/min	75 l/min	100 l/min
Dew Point	-11°C / 12°F				
Particles	<0.01µm				
Phthalates	None				
Suspended Liquids	None				
Pressure Gauges	3				
Start-Up Time	30 minutes				

General

	1031	Hi-Flow	1032	1033	1034
Dimensions (cm)	37.1x20x76	37.1x20x76	42.8x25x73	42.8x25x73	36.9x25x96
Dimensions (ins)	14.5x7.9x29.9	14.5x7.9x29.9	16.8x9.8x28.7	16.8x9.8x28.7	14.5x9.8x37.8
Generator Weight	14/30.8	14/30.8	18/39.6	19.5/42.9	21/46.2
Shipping Weight	17.5/38.5	17.5/38.5	21.5/47.3	23/50.6	25/55

Introduction

The Peak Scientific Nitrogen Generator is designed specifically for use with Mass Spectrometer applications. The Generator has been designed to produce Nitrogen from an existing source of dry, oil free, compressed air in the laboratory.

Unpacking and Installation

Although Peak Scientific take every precaution with safe transit and packaging, it is advisable to fully inspect the unit for any sign of transit damage.

Check 'SHOCKWATCH' label for signs of rough handling prior to un-packing –



Any damage should be reported immediately to the carrier and Peak Scientific or the Peak Partner from where the unit was purchased.

Follow the unpacking instructions posted on the side of the crate. It will require two people to remove the unit from the shipping crate and to manoeuvre the Generator onto the floor.

Please save the product packaging for storage or future shipment of the Generator.



After unpacking and a visual inspection, the unit should be placed in a ventilated area away from direct sunlight. Care should be taken not to obstruct the ventilation holes on the side of the unit. The Generator should be placed on a steady and level base. The Infinity 103 has been designed to fit under most workbenches. Alternatively, the unit may be wall mounted using the fixing slots provided.*



Performance of the Generator (like all sophisticated equipment) is affected by ambient temperatures. Prolonged operation in temperatures exceeding 30°C will shorten the life of the unit.

Fittings Kit


Supplied in the Fittings Kit are all the fittings required to connect the Infinity 103* Generator to the application. The contents of the Fittings Kit are as follows:

- | | |
|-------------------------------------|-------|
| 1. ¼" Compression Fittings | × 3 |
| 2. ¼" Teflon Tubing | × 12m |
| 3. Screw N8 x 1 ½ Supa Counter Sunk | × 4 * |
| 4. Raw-Plug for N8 Screw (521-3087) | × 4 * |
| 5. Installation Guide-Infinity 103X | × 1** |
| 6. User Manual CD- Infinity 103X | × 1 |


* Note 1032 & 1033 will have 3 off items 3 and 4.

** Installation guide will be corresponding to specific generator model.


General Dimensions



Infinity 1031/Hi-Flow in mm



Infinity 1032 & 1033 in mm



Infinity 1034 in mm

Air Connection

Infinity 103* Generator should be connected to a clean, dry, OIL - FREE source of compressed air. A minimum inlet pressure of 120 psig (8.3 barg) is required. Any doubts as to the suitability of your compressed air supply should be referred to Peak Scientific or any of their authorised partners.

The generator has a *Breathing Air Filter* with ¼" BSPT connection to the left side of the unit. The Compressed Air supply should be connected here. This filter will drain moisture and is equipped with an automatic drain. The drain should be led to a convenient place. There are 3-off ¼" BSPT Female bulkhead connections to the right of the unit. Fittings are provided to connect these ports to the gas connections of the Mass Spectrometer.

LC/MS with 3-off Connections (e.g. Api2000)

Connect the 3-off gas ports at the right side of the generator to their respective ports on the Mass Spectrometer using the 3-off ¼" BSPT x ¼" Swagelok fittings. Use ¼" tubing throughout.

LC/MS with 4-off Connections (e.g. Api3000)

Connect *Curtain Gas* and *Exhaust* to their respective ports using ¼" tubing as per above. The *Source Gas* will have to be split into *Gases 1 & 2* at the Mass Spectrometer. Connect *Gases 1 & 2* with 1/8" tubing to the ¼" Tee Piece with the 2-off ¼" BSPT x 1/8" Swagelok fittings. Connect the inlet branch of the tee to the *Source Gas* connection at the rear of the generator using ¼" tubing as above.

NB: - To minimize pressure losses position the Tee as close to the Mass Spectrometer as possible

Commissioning

With the Generator installed (as described earlier), connect compressed air to the unit and open the air supply to the Unit. Disconnect the Outlet connections to allow the generator to vent to atmosphere until the unit is stabilised. The Generator has been pre-set in the factory to give the specified output flow-rate and pressure in line with known requirements. Once the Membranes reach the design pressures the Generator will stabilise and produce Nitrogen. Maximum purity will be achieved after approximately 30 minutes. The generator can then be re-connected to the application.

The design of the generator is such that it can deliver up to the supplies specified in the *Pressure & Flow Settings* table below. Should the demand for gasses be less than the rated output flow at any time, the demand will be determined by the consuming equipment. In circumstances of no demand, the Generator remains operational and will provide Nitrogen immediately as demand resumes.

Pressure & Flow Settings

Port	Usage/Connection	Flow/Pressure (l/min @ psi)				
		1031	Hi-Flow	1032	1033	1034
Curtain	Curtain Gas	27 @ 80	12 @ 80	52 @ 80	85 @ 80	110 @ 80
Source	Ion Source/Nebuliser	46 @ 110	26 @ 110	55 @ 110	85 @ 110	104 @ 110
Exhaust	Exhaust Gas	23 @ 70	16 @ 60	45 @ 70	60 @ 80	100 @ 70

The above settings should allow the Infinity 103* to be operated with all standard configurations for the AB Sciex API range of LC/MS/MS instruments. Should the above settings not provide sufficient flow or pressure for your application please contact Peak Scientific for assistance.

Cleaning

Clean the outside of the Generator only using warm soapy water and a clean damp cloth. Ensure the cloth is thoroughly rung out to remove excess fluid prior to use.



Under no circumstances should any solvents or abrasive cleaning solutions be used as these can contain fumes that could be harmful to the Generator.

Maintenance



*Servicing and/or repair of the Generator should only be undertaken by a **TECHNICALLY COMPETENT PERSON** with the Generator in a safely isolated condition.*

Due to the simplicity of the design and the small number of moving parts, the Infinity 103* Nitrogen Generator will have a long and trouble free life. However as with all scientific and technical equipment it should be regularly inspected and serviced as below.

Your Infinity 103* Nitrogen Generator requires a service annually. A service kit is available from your service representative which includes all the consumables required to maintain your generator. The kit also offers a 10% discount over purchasing the components individually.

Service Schedule

Service interval	Component	Visit	Qty.
12 months	Infinity 1030 Series Annual Service Kit	www.peakscientific.com/ordering	1

Safely Isolated Condition

The unit is in a safely isolated condition when it is disconnected from its application and fully de-pressurised. Directions for isolating the Generator are shown below.



Failure to place the Generator in a safely isolated condition when instructed to do so may lead to personal injury or injury to others and even death.

- Disconnect from air supply.
- Ensure the output pressure gauge read zero. (If gauge does not fall to zero, loosen the outlet fitting slightly to allow trapped gas to escape).

Disconnect from the application.

IMPORTANT DOCUMENTS

Warranty Entitlement

To register your generator for your warranty entitlement, send the completed form to Peak Scientific by:

- *Email* warranty@peakscientific.com
- *Online* http://www.peakscientific.com/service-and-support/warranty_registration
- *Phone* +44 (0)141 530 4185
- *Fax* +44 (0)141 812 8200

PRODUCT WARRANTY REGISTRATION	
COMPANY:	CONTACT NAME:
ADDRESS:	
	EMAIL ADDRESS:
CITY/TOWN:	GENERATOR SERIAL NUMBER:
POSTCODE:	
COUNTRY:	MODEL TYPE:
TELEPHONE:	INSTALLATION DATE (DD/MM/YYYY):

Important Please Note:

You have 1 month to register your Peak Scientific product from the date of shipment.

If you wish to defer installation of your generator you must notify Peak Scientific within 1 month of the shipment date. This can be done by emailing warranty@peakscientific.com Once registered the warranty will be honoured for a period of 12 months after the installation date.

For any generators that remain unregistered the warranty will begin from date of shipment.

Thank you on behalf of Peak Scientific.

Introduction


The following is a guide for maintenance of the Infinity 103x Nitrogen Generator as provided for by the Peak Service Plan.

This procedure is to be performed on an annual basis.

1. Check that the LC/MS is not in use and that it is OK to disconnect the Gas Supply. Check the Gas Demand is turned OFF.
2. Shut the Air Supply Valve.



3. Remove the Front Access Panel.
4. Carefully slacken the Gas Outlet Lines from the right side of the unit and allow it to depressurize. Confirm that the gauges return to Zero. **Any pressure gauge that does not return to zero may be damaged and must be replaced.**
5. Replace the 2-off RAC Filter Elements.



6. Unscrew the top and bottom covers from the breathing air filter.



7. Replace the Breathing Air Filter Elements and re-assemble breathing air filter.
8. Carefully open the Air Supply Valve.
9. Re-tighten the Nitrogen Outlet Lines and re-fit the Front Panel.
10. Switch the LC/MS gas demand ON.

Service Parts List

Part Number	Description	Quantity
3302277 (00-0039*)	RAC Filter Element	2
3302280 (00-0045*)	Coalescing Element for BA Filter	1
3302281 (00-0046*)	Active Carbon Element for BA Filter	1

Element part numbers with (*) are obsolete, the 330XXXX part number should be ordered.

Peak Scientific UK
Fountain Crescent
Inchinnan Business Park
Inchinnan
PA4 9RE
Scotland, UK

Tel: +44 (0)141 812 8100
Fax: +44 (0)141 812 8200

Peak Scientific Germany
Herriotstrasse 1
60528 Frankfurt

Germany

Tel: +49 (0)69 677 33 490
Fax: +49 (0)69 677 33 200

Peak Scientific Taiwan
4F.-6, No.736
Zhongzheng Rd.
Zhonghe Dist.
New Taipei City

Tel: +886 2 8226 2383
Fax: +886 2 8226 9499

Peak Scientific North America
19 Sterling Road
Suite #1
Billerica, MA 01862

USA

Tel: +1 866 647 1649
Fax: +1 978 608 9503

Peak Scientific China
Room 606 Building 1
Lane 2277 Zuchongzhi Road
Pudong New Area
Shanghai
201203, China

Tel: +86 21 5079 1190
Fax: +86 21 5079 1191

Peak Scientific Brazil
Peak Scientific Brasil
Av. Quieroz Filho
1700 - torre Sunny-Conj. 602
São Paulo SP Brasil

Tel: +1 866 647 1649

Peak Scientific India
202, Amsri Shamira
Old Lancer Line
Opp. St. Mary's Degree
College
S.D. Road Secunderabad
500 025, India

Tel: +91 40 2780 0663
Fax: +91 40 2780 0663

Peak Scientific Mexico
Solon 352
Col. Los Morales Polanco
11530 Mexico, D.F.

Mexico

Tel: +1 866 647 1649
Fax: +1 978 608 9503

Peak Scientific Japan K.K.
2-7-56, 2F Fuji Building 28
Kita Aoyama, Minato-Ku
Tokyo, Japan
107-0061

Tel: +81 3-6864-0468

Peak Scientific Africa
PO Box 478
Somerset Mall
7137

South Africa

Tel: +27 (0)2185 16542
Fax: +27 (0)8654 64473

Peak Scientific Australia
PO Box 65
Belmont
Victoria 3216

Australia

Tel: +61 1300 965 352

Peak Scientific Singapore
3 Science Park Drive
#03-14 The Franklin
Science Park Drive 1
Singapore, 118223

Tel: +65-6777 1966
Fax: +65-6777 1978